

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064831.D
 Acq On : 25 Nov 2020 00:18
 Operator : JC/MD
 Sample : L4798-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 22BR-20201118

Quant Time: Nov 25 06:23:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	165708	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	294430	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	304469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	138746	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102366	53.14	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	106.28%		
35) Dibromofluoromethane	7.55	113	92500	52.50	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	105.00%		
50) Toluene-d8	10.06	98	363975	50.56	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.12%		
62) 4-Bromofluorobenzene	12.37	95	143510	55.71	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.42%		
Target Compounds						
27) cis-1,2-Dichloroethene	6.78	96	30979	15.428	ug/l	89
44) Trichloroethene	8.80	130	21442	8.690	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064831.D
Acq On : 25 Nov 2020 00:18
Operator : JC/MD
Sample : L4798-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
22BR-20201118

Quant Time: Nov 25 06:23:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

