

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064834.D
 Acq On : 25 Nov 2020 1:30
 Operator : JC/MD
 Sample : L4798-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 48BR-20201118

Quant Time: Nov 25 06:31:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	156801	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	273975	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	285566	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	131981	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	97436	53.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.90%
35) Dibromofluoromethane	7.55	113	89341	54.50	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.00%
50) Toluene-d8	10.06	98	345648	51.60	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.20%
62) 4-Bromofluorobenzene	12.37	95	133517	55.70	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.40%

Target Compounds

					Qvalue
4) Vinyl Chloride	2.16	62	1368	0.743	ug/l 92
9) 1,1,2-Trichlorotrifluoroet	3.71	101	2687	1.760	ug/l # 86
12) 1,1-Dichloroethene	3.69	96	9047	5.950	ug/l 89
27) cis-1,2-Dichloroethene	6.78	96	39951	21.027	ug/l 87
44) Trichloroethene	8.80	130	1168492	508.914	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064834.D
Acq On : 25 Nov 2020 1:30
Operator : JC/MD
Sample : L4798-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
48BR-20201118

Quant Time: Nov 25 06:31:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

