

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064835.D
 Acq On : 25 Nov 2020 1:54
 Operator : JC/MD
 Sample : L4798-07DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 48BR-20201118DL

Quant Time: Nov 25 06:34:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	158529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	278191	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	293187	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	131996	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	99461	53.97	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.94%	
35) Dibromofluoromethane		7.55	113	90432	54.33	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.66%	
50) Toluene-d8		10.06	98	350636	51.55	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.10%	
62) 4-Bromofluorobenzene		12.38	95	135314	55.59	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.17	62	534	0.287	ug/l #	80
12) 1,1-Dichloroethene	3.70	96	1007	0.655	ug/l #	62
27) cis-1,2-Dichloroethene	6.78	96	6417	3.341	ug/l	92
44) Trichloroethene	8.80	130	124030	53.200	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064835.D
Acq On : 25 Nov 2020 1:54
Operator : JC/MD
Sample : L4798-07DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
48BR-20201118DL

Quant Time: Nov 25 06:34:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

