

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064837.D
 Acq On : 25 Nov 2020 2:42
 Operator : JC/MD
 Sample : L4798-11DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 91BR-20201118DL

Quant Time: Nov 25 06:38:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	161825	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	281394	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	287595	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	128997	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	97387	51.77	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.54%
35) Dibromofluoromethane	7.55	113	90564	53.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.58%
50) Toluene-d8	10.06	98	346906	50.42	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.84%
62) 4-Bromofluorobenzene	12.38	95	133475	54.21	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.42%

Target Compounds

					Qvalue
4) Vinyl Chloride	2.17	62	20840	10.968	ug/l 97
27) cis-1,2-Dichloroethene	6.78	96	91716	46.773	ug/l 89
44) Trichloroethene	8.80	130	2230	0.946	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064837.D
Acq On : 25 Nov 2020 2:42
Operator : JC/MD
Sample : L4798-11DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
91BR-20201118DL

Quant Time: Nov 25 06:38:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

