

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064840.D
 Acq On : 25 Nov 2020 3:54
 Operator : JC/MD
 Sample : L4798-08 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 56BR-20201118

Quant Time: Nov 25 06:43:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	160487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	282824	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	293781	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	128951	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	100208	53.71	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	107.42%		
35) Dibromofluoromethane	7.55	113	90975	53.76	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	107.52%		
50) Toluene-d8	10.06	98	356573	51.57	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.14%		
62) 4-Bromofluorobenzene	12.37	95	133011	53.75	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	107.50%		

Target Compounds

					Qvalue
4) Vinyl Chloride	2.16	62	2450	1.300	ug/l 96
27) cis-1,2-Dichloroethene	6.78	96	34938	17.966	ug/l 90
44) Trichloroethene	8.80	130	222011	93.667	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064840.D
Acq On : 25 Nov 2020 3:54
Operator : JC/MD
Sample : L4798-08 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
56BR-20201118

Quant Time: Nov 25 06:43:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

