

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN112420\
 Data File : VN064822.D
 Acq On : 24 Nov 2020 19:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 25 05:36:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.433	0.393	9.2	95	0.00
3 P	Chloromethane	0.563	0.496	11.9	93	0.00
4 C	Vinyl Chloride	0.587	0.543	7.5#	97	0.00
5 T	Bromomethane	0.385	0.287	25.5#	78	0.00
6 T	Chloroethane	0.363	0.341	6.1	97	0.00
7 T	Trichlorofluoromethane	0.812	0.759	6.5	97	0.00
8 T	Diethyl Ether	0.320	0.310	3.1	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.487	0.457	6.2	96	0.00
10 T	Methyl Iodide	0.631	0.336	46.8#	53	0.00
11 T	Tert butyl alcohol	0.091	0.082	9.9	89	-0.01
12 CM	1,1-Dichloroethene	0.485	0.464	4.3#	99	0.00
13 T	Acrolein	0.055	0.040	27.3#	73	0.00
14 T	Allyl chloride	0.800	0.755	5.6	96	0.00
15 T	Acrylonitrile	0.250	0.241	3.6	94	0.00
16 T	Acetone	0.191	0.187	2.1	98	0.00
17 T	Carbon Disulfide	1.397	1.196	14.4	97	0.00
18 T	Methyl Acetate	0.575	0.547	4.9	97	0.00
19 T	Methyl tert-butyl Ether	1.555	1.560	-0.3	98	0.00
20 T	Methylene Chloride	0.599	0.547	8.7	100	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.501	6.9	97	0.00
22 T	Diisopropyl ether	1.601	1.615	-0.9	98	0.00
23 T	Vinyl Acetate	1.280	1.320	-3.1	98	0.00
24 P	1,1-Dichloroethane	0.957	0.941	1.7	98	0.00
25 T	2-Butanone	0.310	0.310	0.0	94	0.00
26 T	2,2-Dichloropropane	0.830	0.728	12.3	89	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.585	3.5	98	0.00
28 T	Bromochloromethane	0.438	0.427	2.5	99	0.00
29 T	Tetrahydrofuran	0.214	0.208	2.8	93	0.00
30 C	Chloroform	0.986	0.956	3.0#	99	0.00
31 T	Cyclohexane	0.810	0.766	5.4	100	0.00
32 T	1,1,1-Trichloroethane	0.847	0.813	4.0	97	0.00
33 S	1,2-Dichloroethane-d4	0.581	0.567	2.4	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.299	0.308	-3.0	108	0.00
36 T	1,1-Dichloropropene	0.463	0.449	3.0	99	0.00
37 T	Ethyl Acetate	0.446	0.413	7.4	94	0.00
38 T	Carbon Tetrachloride	0.482	0.454	5.8	96	0.00
39 T	Methylcyclohexane	0.499	0.494	1.0	102	0.00
40 TM	Benzene	1.420	1.345	5.3	98	0.00
41 T	Methacrylonitrile	0.206	0.199	3.4	109	0.00
42 TM	1,2-Dichloroethane	0.479	0.440	8.1	96	0.00
43 T	Isopropyl Acetate	0.718	0.668	7.0	93	0.00
44 TM	Trichloroethene	0.419	0.380	9.3	95	0.00
45 C	1,2-Dichloropropane	0.368	0.361	1.9#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN112420\
 Data File : VN064822.D
 Acq On : 24 Nov 2020 19:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 25 05:36:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.246	0.236	4.1	99	0.00
47 T	Bromodichloromethane	0.494	0.474	4.0	98	0.00
48 T	Methyl methacrylate	0.343	0.333	2.9	95	0.00
49 T	1,4-Dioxane	0.006	0.006	0.0	91	0.00
50 S	Toluene-d8	1.222	1.199	1.9	103	0.00
51 T	4-Methyl-2-Pentanone	0.431	0.416	3.5	94	0.00
52 CM	Toluene	0.875	0.844	3.5#	98	0.00
53 T	t-1,3-Dichloropropene	0.551	0.523	5.1	97	0.00
54 T	cis-1,3-Dichloropropene	0.590	0.570	3.4	97	0.00
55 T	1,1,2-Trichloroethane	0.360	0.342	5.0	100	0.00
56 T	Ethyl methacrylate	0.496	0.501	-1.0	97	0.00
57 T	1,3-Dichloropropane	0.593	0.572	3.5	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.276	-1.1	96	0.00
59 T	2-Hexanone	0.310	0.302	2.6	94	0.00
60 T	Dibromochloromethane	0.400	0.383	4.3	96	0.00
61 T	1,2-Dibromoethane	0.365	0.353	3.3	98	0.00
62 S	4-Bromofluorobenzene	0.437	0.429	1.8	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.486	0.400	17.7	86	0.00
65 PM	Chlorobenzene	1.074	1.026	4.5	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.395	0.381	3.5	97	0.00
67 C	Ethyl Benzene	1.794	1.757	2.1#	98	0.00
68 T	m/p-Xylenes	0.679	0.680	-0.1	99	0.00
69 T	o-Xylene	0.644	0.647	-0.5	99	0.00
70 T	Styrene	1.088	1.127	-3.6	100	0.00
71 P	Bromoform	0.315	0.293	7.0	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.478	3.576	-2.8	100	0.00
74 T	N-amyl acetate	1.369	1.344	1.8	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.019	1.019	0.0	102	0.00
76 T	1,2,3-Trichloropropane	1.077	0.995	7.6	92	0.00
77 T	Bromobenzene	0.938	0.894	4.7	98	0.00
78 T	n-propylbenzene	3.962	3.957	0.1	97	0.00
79 T	2-Chlorotoluene	2.449	2.390	2.4	98	0.00
80 T	1,3,5-Trimethylbenzene	2.891	3.025	-4.6	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.393	0.348	11.5	85	0.00
82 T	4-Chlorotoluene	2.567	2.475	3.6	97	0.00
83 T	tert-Butylbenzene	2.423	2.568	-6.0	103	0.00
84 T	1,2,4-Trimethylbenzene	2.923	3.037	-3.9	100	0.00
85 T	sec-Butylbenzene	3.217	3.317	-3.1	100	0.00
86 T	p-Isopropyltoluene	2.929	3.100	-5.8	100	0.00
87 T	1,3-Dichlorobenzene	1.666	1.598	4.1	99	0.00
88 T	1,4-Dichlorobenzene	1.762	1.574	10.7	96	0.00
89 T	n-Butylbenzene	2.572	2.551	0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN112420\
Data File : VN064822.D
Acq On : 24 Nov 2020 19:53
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 25 05:36:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.509	0.522	-2.6	99	0.00
91 T	1,2-Dichlorobenzene	1.636	1.555	5.0	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.229	0.220	3.9	98	0.00
93 T	1,2,4-Trichlorobenzene	1.012	0.913	9.8	95	0.00
94 T	Hexachlorobutadiene	0.506	0.420	17.0	96	0.00
95 T	Naphthalene	3.031	2.834	6.5	96	0.00
96 T	1,2,3-Trichlorobenzene	0.994	0.900	9.5	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6