

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\  
 Data File : VN064814.D  
 Acq On : 24 Nov 2020 16:40  
 Operator : JC/MD  
 Sample : L4829-16  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-5

Manual Integrations  
 APPROVED

MMDadoda  
 11/25/2020 3:57:29 PM

Quant Time: Nov 25 05:14:01 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N112320W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Nov 23 13:54:12 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 7.63  | 168  | 174923   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.55  | 114  | 314191   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.38 | 117  | 320159   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.32 | 152  | 156112   | 50.00 | ug/l  | 0.00     |

## System Monitoring Compounds

|                           |        |       |          |          |      |         |
|---------------------------|--------|-------|----------|----------|------|---------|
| 33) 1,2-Dichloroethane-d4 | 7.99   | 65    | 103045   | 50.67    | ug/l | 0.00    |
| Spiked Amount             | 50.000 | Range | 61 - 141 | Recovery | =    | 101.34% |
| 35) Dibromofluoromethane  | 7.55   | 113   | 94517    | 50.27    | ug/l | 0.00    |
| Spiked Amount             | 50.000 | Range | 69 - 133 | Recovery | =    | 100.54% |
| 50) Toluene-d8            | 10.06  | 98    | 394551   | 51.36    | ug/l | 0.00    |
| Spiked Amount             | 50.000 | Range | 65 - 126 | Recovery | =    | 102.72% |
| 62) 4-Bromofluorobenzene  | 12.37  | 95    | 158500   | 57.66    | ug/l | 0.00    |
| Spiked Amount             | 50.000 | Range | 58 - 135 | Recovery | =    | 115.32% |

## Target Compounds

|                             | R.T.  | QIon | Response | Conc   | Units  | Qvalue |
|-----------------------------|-------|------|----------|--------|--------|--------|
| 16) Acetone                 | 3.78  | 43   | 14139    | 21.150 | ug/l   | 93     |
| 19) Methyl tert-butyl Ether | 5.00  | 73   | 130645   | 24.015 | ug/l # | 63     |
| 25) 2-Butanone              | 6.80  | 43   | 3166m    | 2.922  | ug/l   |        |
| 31) Cyclohexane             | 7.61  | 56   | 49187    | 17.354 | ug/l # | 95     |
| 39) Methylcyclohexane       | 9.05  | 83   | 83992    | 26.791 | ug/l   | 99     |
| 40) Benzene                 | 8.00  | 78   | 237862   | 26.654 | ug/l   | 100    |
| 52) Toluene                 | 10.12 | 92   | 5997     | 1.091  | ug/l   | 94     |
| 64) Tetrachloroethene       | 10.59 | 164  | 1122     | 0.360  | ug/l # | 84     |
| 67) Ethyl Benzene           | 11.48 | 91   | 7228     | 0.629  | ug/l   | 96     |
| 68) m/p-Xylenes             | 11.59 | 106  | 32336    | 7.436  | ug/l   | 98     |
| 69) o-Xylene                | 11.92 | 106  | 9714     | 2.354  | ug/l   | 91     |
| 73) Isopropylbenzene        | 12.22 | 105  | 34514    | 3.178  | ug/l   | 99     |
| 78) n-propylbenzene         | 12.56 | 91   | 16494    | 1.333  | ug/l   | 98     |
| 80) 1,3,5-Trimethylbenzene  | 12.71 | 105  | 16486    | 1.827  | ug/l   | 98     |
| 83) tert-Butylbenzene       | 12.96 | 119  | 3026     | 0.400  | ug/l   | 90     |
| 84) 1,2,4-Trimethylbenzene  | 13.01 | 105  | 30784    | 3.373  | ug/l   | 98     |
| 85) sec-Butylbenzene        | 13.14 | 105  | 19782    | 1.969  | ug/l # | 73     |
| 86) p-Isopropyltoluene      | 13.26 | 119  | 5027     | 0.550  | ug/l # | 51     |
| 89) n-Butylbenzene          | 13.56 | 91   | 18109    | 2.255  | ug/l # | 1      |
| 95) Naphthalene             | 15.10 | 128  | 64911    | 6.858  | ug/l   | 99     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\  
Data File : VN064814.D  
Acq On : 24 Nov 2020 16:40  
Operator : JC/MD  
Sample : L4829-16  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 18 Sample Multiplier: 1


Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-5

Manual Integrations  
APPROVED

MMDadoda  
11/25/2020 3:57:29 PM

Quant Time: Nov 25 05:14:01 2020  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N112320W.M  
Quant Title : SW846 8260  
QLast Update : Mon Nov 23 13:54:12 2020  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN064814.D

Acq: 24 Nov 2020 16:40

Instrument :

MSVOA\_N

ClientSampled :

MW-5

Tot Ion:168 Resp: 174923

Ion Ratio Lower Upper

168 100


99 54.9 43.0 64.4

Manual Integrations

APPROVED

MMDadoda

11/25/2020 3:57:29 PM



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

80000



60000

40000

20000

0

Time-->



#16

Acetone

Concen: 21.150 ug/l

RT: 3.78 min Scan# 634

Delta R.T. 0.00 min

Lab File: VN064814.D


Acq: 24 Nov 2020 16:40

Tgt Ion: 43 Resp: 14139

Ion Ratio Lower Upper

43 100

58 36.0 25.7 38.5



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO

3.78

5000

4000


3000


2000

1000

0

Time-->









#19  
Methvl tert-butvl Ether  
Concen: 24.015 ug/l  
RT: 5.00 min Scan# 1014  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampled :  
MW-5

Tot Ion: 73 Resp: 130645  
Ion Ratio Lower Upper  
73 100  
57 39.0 17.4 26.0#




Manual Integrations  
APPROVED


MMDadoda  
11/25/2020 3:57:29 PM



#25  
2-Butanone  
Concen: 2.922 ug/l m  
RT: 6.80 min Scan# 1575  
Delta R.T. 0.01 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Tgt Ion: 43 Resp: 3166  
Ion Ratio Lower Upper  
43 100  
72 25.1 21.0 31.6









#31  
Cyclohexane  
Concen: 17.354 ug/l  
RT: 7.61 min Scan# 1826  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-5

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 56      | 100   |       |       |
| 69      | 42.9  | 27.5  | 41.3# |
| 84      | 93.9  | 73.9  | 110.9 |




Manual Integrations  
APPROVED


MMDadoda  
11/25/2020 3:57:29 PM



#33  
1,2-Dichloroethane-d4  
Concen: 50.671 ug/l  
RT: 7.99 min Scan# 1943  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 65      | 100   |       |       |
| 67      | 54.6  | 0.0   | 108.2 |









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.55 min Scan# 2118  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-5

| Tot Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 114     | 100   | 314191 |       |       |
| 63      | 19.0  | 0.0    | 38.6  |       |
| 88      | 15.7  | 0.0    | 32.6  |       |




Manual Integrations  
APPROVED


MMDadoda  
11/25/2020 3:57:29 PM



#35  
Dibromofluoromethane  
Concen: 50.275 ug/l  
RT: 7.55 min Scan# 1807  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

| Tgt Ion | Ratio | Resp  | Lower | Upper |
|---------|-------|-------|-------|-------|
| 113     | 100   | 94517 |       |       |
| 111     | 103.1 | 81.6  | 122.4 |       |
| 192     | 19.6  | 16.2  | 24.2  |       |









#39  
Methvlcvclohexane  
Concen: 26.791 ug/l  
RT: 9.05 min Scan# 2272  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampled :  
MW-5

| Tot Ion   | 83   | Resp  | 83992 |
|-----------|------|-------|-------|
| Ion Ratio | 100  | Lower | Upper |
| 83        | 100  |       |       |
| 55        | 76.7 | 60.2  | 90.2  |
| 98        | 46.6 | 37.4  | 56.2  |




Manual Integrations  
APPROVED


MMDadoda  
11/25/2020 3:57:29 PM



#40  
Benzene  
Concen: 26.654 ug/l  
RT: 8.00 min Scan# 1948  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

| Tgt Ion   | 78   | Resp  | 237862 |
|-----------|------|-------|--------|
| Ion Ratio | 100  | Lower | Upper  |
| 78        | 100  |       |        |
| 77        | 22.2 | 17.9  | 26.9   |





#50

Toluene-d8

Concen: 51.363 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN064814.D

Acq: 24 Nov 2020 16:40

Instrument :

MSVOA\_N

ClientSampled :

MW-5

Tot Ion: 98 Resp: 394551

Ion Ratio Lower Upper

98 100


100 63.5 51.7 77.5

Manual Integrations

APPROVED

MMDadoda

11/25/2020 3:57:29 PM



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN

250000

200000

150000

100000

50000



0

10.06

10.10

10.20

Time-->



#52

Toluene

Concen: 1.091 ug/l

RT: 10.12 min Scan# 2607

Delta R.T. -0.00 min

Lab File: VN064814.D


Acq: 24 Nov 2020 16:40

Tgt Ion: 92 Resp: 5997

Ion Ratio Lower Upper

92 100

91 182.9 139.1 208.7



Abundance Ion 92.00 (91.70 to 92.70): VN0

Ion 91.00 (90.70 to 91.70): VN0

6000

5000

4000

3000

2000


1000


0

10.12

10.15

Time-->









#62  
4-Bromofluorobenzene  
Concen: 57.655 ug/l  
RT: 12.37 min Scan# 3307  
Delta R.T. -0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-5

| Tot Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 95      | 100   | 158500 |       |       |
| 174     | 79.9  | 0.0    | 160.4 |       |
| 176     | 77.4  | 0.0    | 158.4 |       |




Manual Integrations  
APPROVED


MMDadoda  
11/25/2020 3:57:29 PM



#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.38 min Scan# 2998  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

| Tgt Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 117     | 100   | 320159 |       |       |
| 82      | 53.7  | 44.2   | 66.2  |       |
| 119     | 32.3  | 25.8   | 38.6  |       |









#64  
Tetrachloroethene  
Concen: 0.360 ug/l  
RT: 10.59 min Scan# 2754  
Delta R.T. -0.01 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-5

| Tot Ion | Ratio | Lower | Upper  |
|---------|-------|-------|--------|
| 164     | 100   |       |        |
| 166     | 106.7 | 101.4 | 152.0  |
| 129     | 101.2 | 76.1  | 114.1  |
| 131     | 69.3  | 74.2  | 111.2# |




Manual Integrations  
APPROVED


MMDadoda  
11/25/2020 3:57:29 PM



#67  
Ethyl Benzene  
Concen: 0.629 ug/l  
RT: 11.48 min Scan# 3030  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 91      | 100   |       |       |
| 106     | 29.3  | 25.2  | 37.8  |






#68  
m/p-Xylenes  
Concen: 7.436 ug/l  
RT: 11.59 min Scan# 3065  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampled :  
MW-5




Tot Ion:106 Resp: 32336  
Ion Ratio Lower Upper  
106 100  
91 198.8 161.0 241.4

Manual Integrations  
APPROVED

MMDadoda  
11/25/2020 3:57:29 PM




Abundance Ion 106.00 (105.70 to 106.70): V  
Ion 91.00 (90.70 to 91.70): VN0






#69  
o-Xylene  
Concen: 2.354 ug/l  
RT: 11.92 min Scan# 3165  
Delta R.T. -0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Tgt Ion:106 Resp: 9714  
Ion Ratio Lower Upper  
106 100  
91 224.3 105.5 316.5



Abundance Ion 106.00 (105.70 to 106.70): V  
Ion 91.00 (90.70 to 91.70): VN0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3600

Delta R.T. -0.00 min

Lab File: VN064814.D

Acq: 24 Nov 2020 16:40

Instrument :

MSVOA\_N





ClientSampled :

MW-5

| Tot Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 152     | 100   | 156112 |       |       |
| 115     | 60.5  | 29.7   | 89.1  |       |
| 150     | 155.8 | 0.0    | 350.6 |       |

Manual Integrations  
APPROVED

MMDadoda  
11/25/2020 3:57:29 PM



#73

Isopropylbenzene

Concen: 3.178 ug/l




RT: 12.22 min Scan# 3260


Delta R.T. 0.00 min

Lab File: VN064814.D

Acq: 24 Nov 2020 16:40

| Tgt Ion | Ratio | Resp  | Lower | Upper |
|---------|-------|-------|-------|-------|
| 105     | 100   | 34514 |       |       |
| 120     | 27.1  | 13.3  | 39.9  |       |





#78

n-propylbenzene

Concen: 1.333 ug/l

RT: 12.56 min Scan# 3366

Delta R.T. -0.00 min

Lab File: VN064814.D

Acq: 24 Nov 2020 16:40

Instrument :

MSVOA\_N

ClientSampled :

MW-5

Tot Ion: 91 Resp: 16494

Ion Ratio Lower Upper

91 100


120 24.3 11.6 34.7

Manual Integrations

APPROVED




MMDadoda

11/25/2020 3:57:29 PM



Abundance Ion 91.00 (90.70 to 91.70): VN064779.D (-3355) (-)

Ion 120.00 (119.70 to 120.70): VN064779.D (-3355) (-)



#80

1,3,5-Trimethylbenzene

Concen: 1.827 ug/l

RT: 12.71 min Scan# 3411

Delta R.T. 0.00 min

Lab File: VN064814.D


Acq: 24 Nov 2020 16:40

Tgt Ion: 105 Resp: 16486

Ion Ratio Lower Upper



105 100


120 50.4 24.6 73.6



Abundance Ion 105.00 (104.70 to 105.70): VN064779.D (-3397) (-)

Ion 120.00 (119.70 to 120.70): VN064779.D (-3397) (-)









#83  
tert-Butylbenzene  
Concen: 0.400 ug/l  
RT: 12.96 min Scan# 3491  
Delta R.T. -0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampled :  
MW-5

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 119     | 100   |       |       |
| 91      | 73.5  | 32.3  | 96.8  |
| 134     | 21.3  | 12.6  | 37.6  |




Manual Integrations  
APPROVED


MMDadoda  
11/25/2020 3:57:29 PM



#84  
1,2,4-Trimethylbenzene  
Concen: 3.373 ug/l  
RT: 13.01 min Scan# 3506  
Delta R.T. 0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 105     | 100   |       |       |
| 120     | 44.8  | 23.1  | 69.2  |





#85

sec-Butylbenzene

Concen: 1.969 ug/l

RT: 13.14 min Scan# 3547

Delta R.T. -0.00 min

Lab File: VN064814.D

Acq: 24 Nov 2020 16:40

Instrument :

MSVOA\_N

ClientSampled :

MW-5

Tot Ion:105 Resp: 19782

Ion Ratio Lower Upper

105 100





134 32.5 10.1 30.1#

Manual Integrations

APPROVED

MMDadoda

11/25/2020 3:57:29 PM



#86

p-Isopropyltoluene

Concen: 0.550 ug/l

RT: 13.26 min Scan# 3584

Delta R.T. 0.00 min

Lab File: VN064814.D

Acq: 24 Nov 2020 16:40




Tgt Ion:119 Resp: 5027


Ion Ratio Lower Upper

119 100

134 0.0 13.1 39.1#

91 0.0 11.5 34.5#









#89  
n-Butylbenzene  
Concen: 2.255 ug/l  
RT: 13.56 min Scan# 3677  
Delta R.T. -0.03 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

Instrument :  
MSVOA\_N  
ClientSampled :  
MW-5

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 91      | 100   |       |       |
| 92      | 0.0   | 26.0  | 78.0# |
| 134     | 123.8 | 12.9  | 38.6# |

Manual Integrations  
APPROVED

MMDadoda  
11/25/2020 3:57:29 PM



#95  
Naphthalene  
Concen: 6.858 ug/l  
RT: 15.10 min Scan# 4155  
Delta R.T. -0.00 min  
Lab File: VN064814.D  
Acq: 24 Nov 2020 16:40

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 128     | 100   |       |       |
| 127     | 12.8  | 10.3  | 15.5  |
| 129     | 10.2  | 8.7   | 13.1  |

