

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064820.D
 Acq On : 24 Nov 2020 19:04
 Operator : JC/MD
 Sample : L4798-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-20201118

Quant Time: Nov 25 13:54:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	172557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	307293	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	316781	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	147367	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	104063	51.87	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.74%	
35) Dibromofluoromethane		7.55	113	95462	51.92	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.84%	
50) Toluene-d8		10.06	98	383287	51.02	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.04%	
62) 4-Bromofluorobenzene		12.37	95	150059	55.81	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.62%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064820.D
Acq On : 24 Nov 2020 19:04
Operator : JC/MD
Sample : L4798-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-20201118

Quant Time: Nov 25 13:54:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN064820.D

Acq: 24 Nov 2020 19:04

Instrument :

MSVOA_N

ClientSampleId :


TB01-20201118

Tot Ion:168 Resp: 172557

Ion Ratio Lower Upper

168 100

99 54.5 43.0 64.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

80000

60000



40000

20000

0

Time-->

7.55 7.60 7.65 7.70



#33

1,2-Dichloroethane-d4

Concen: 51.874 ug/l

RT: 7.99 min Scan# 1943

Delta R.T. 0.00 min

Lab File: VN064820.D


Acq: 24 Nov 2020 19:04

Tgt Ion: 65 Resp: 104063

Ion Ratio Lower Upper

65 100

67 54.1 0.0 108.2



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

7.99

50000

40000

30000


20000


10000

0

Time-->

7.90 8.00 8.10









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2118
 Delta R.T. 0.00 min
 Lab File: VN064820.D
 Acq: 24 Nov 2020 19:04

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-20201118

Tot Ion	Ratio	Lower	Upper
114	100		
63	19.0	0.0	38.6
88	15.3	0.0	32.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 51.917 ug/l
 RT: 7.55 min Scan# 1807
 Delta R.T. 0.00 min
 Lab File: VN064820.D
 Acq: 24 Nov 2020 19:04

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.9	81.6	122.4
192	19.3	16.2	24.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.017 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN064820.D

Acq: 24 Nov 2020 19:04

Instrument :

MSVOA_N

ClientSampleId :


TB01-20201118

Tot Ion: 98 Resp: 383287

Ion Ratio Lower Upper

98 100

100 64.4 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN064779.D (-2572) (-)

Ion 100.00 (99.70 to 100.70): VN064779.D (-2572) (-)

10.06

250000

200000



150000

100000

50000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 55.810 ug/l

RT: 12.37 min Scan# 3307

Delta R.T. -0.00 min

Lab File: VN064820.D

Acq: 24 Nov 2020 19:04


Tgt Ion: 95 Resp: 150059

Ion Ratio Lower Upper

95 100

174 81.8 0.0 160.4

176 79.1 0.0 158.4



Abundance Ion 94.90 (94.60 to 95.60): VN064779.D (-3294) (-)

Ion 173.80 (173.50 to 174.50): VN064779.D (-3294) (-)

Ion 175.80 (175.50 to 176.50): VN064779.D (-3294) (-)

12.37

120000

100000

80000


60000


40000

20000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN064820.D
Acq: 24 Nov 2020 19:04

Instrument :
MSVOA_N
ClientSampleId :
TB01-20201118

Tot Ion:117 Resp: 316781
Ion Ratio Lower Upper
117 100
82 55.5 44.2 66.2
119 32.6 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3600
Delta R.T. -0.00 min
Lab File: VN064820.D
Acq: 24 Nov 2020 19:04

Tgt Ion:152 Resp: 147367
Ion Ratio Lower Upper
152 100
115 60.1 29.7 89.1
150 158.7 0.0 350.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

