

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112421\
 Data File : VN069638.D
 Acq On : 24 Nov 2021 13:25
 Operator : JC/MD
 Sample : M4815-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Nov 26 06:57:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1400905	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	2468480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2473352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	925893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1063708	52.367	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.740%	
35) Dibromofluoromethane	8.021	113	724959	48.968	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.940%	
50) Toluene-d8	10.443	98	3123011	53.635	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.260%	
62) 4-Bromofluorobenzene	12.731	95	1125397	52.612	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.220%	
Target Compounds						
16) Acetone	4.293	43	22072	2.014	ug/l	99
52) Toluene	10.505	92	15101	0.378	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112421\
Data File : VN069638.D
Acq On : 24 Nov 2021 13:25
Operator : JC/MD
Sample : M4815-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Nov 26 06:57:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

