

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112421\
 Data File : VN069641.D
 Acq On : 24 Nov 2021 14:41
 Operator : JC/MD
 Sample : M4815-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7

Quant Time: Nov 26 06:59:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1551666	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2695692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2727204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1218101	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1138560	50.606	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 101.220%		
35) Dibromofluoromethane	8.024	113	785943	48.612	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 97.220%		
50) Toluene-d8	10.443	98	3405483	53.556	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 107.120%		
62) 4-Bromofluorobenzene	12.731	95	1297994	55.566	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 111.140%		
Target Compounds						
						Qvalue
16) Acetone	4.301	43	26267	2.164	ug/l	97
95) Naphthalene	15.507	128	68780	5.610	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112421\
Data File : VN069641.D
Acq On : 24 Nov 2021 14:41
Operator : JC/MD
Sample : M4815-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7

Quant Time: Nov 26 06:59:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

