

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112421\
 Data File : VN069643.D
 Acq On : 24 Nov 2021 15:32
 Operator : JC/MD
 Sample : M4815-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Nov 26 06:59:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1616947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2732686	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2766303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1238209	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1141503	48.689	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.380%
35) Dibromofluoromethane	8.024	113	802738	48.979	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.960%
50) Toluene-d8	10.443	98	3471991	53.863	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.720%
62) 4-Bromofluorobenzene	12.731	95	1309400	55.295	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.600%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	49498	3.913	ug/l	91
95) Naphthalene	15.507	128	162663	7.147	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112421\
Data File : VN069643.D
Acq On : 24 Nov 2021 15:32
Operator : JC/MD
Sample : M4815-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Nov 26 06:59:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

