

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112519\
 Data File : VN059381.D
 Acq On : 25 Nov 2019 14:23
 Operator : JC/SP
 Sample : VN1125WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1125WBL01

Quant Time: Nov 25 17:20:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	362380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	552679	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	477689	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	175994	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	204445	47.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.04%	
35) Dibromofluoromethane	7.57	113	164447	49.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.20%	
50) Toluene-d8	10.08	98	660482	50.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.22%	
62) 4-Bromofluorobenzene	12.40	95	211384	45.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112519\
Data File : VN059381.D
Acq On : 25 Nov 2019 14:23
Operator : JC/SP
Sample : VN1125WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1125WBL01

Quant Time: Nov 25 17:20:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration

