

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112519\
 Data File : VN059389.D
 Acq On : 25 Nov 2019 17:38
 Operator : JC/SP
 Sample : K5671-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-112119

Quant Time: Nov 26 01:15:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	363628	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	562047	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	487950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	182922	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	206063	47.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.46%	
35) Dibromofluoromethane	7.57	113	166857	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
50) Toluene-d8	10.08	98	683788	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	12.40	95	217175	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
Target Compounds						
16) Acetone	3.79	43	515969	263.753	ug/l	Qvalue 100
20) Methylene Chloride	4.53	84	160793	36.444	ug/l	99
43) Isopropyl Acetate	8.15	43	81439	8.545	ug/l #	92
95) Naphthalene	15.13	128	11686	4.183	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112519\
Data File : VN059389.D
Acq On : 25 Nov 2019 17:38
Operator : JC/SP
Sample : K5671-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-112119

Quant Time: Nov 26 01:15:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration

