

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064846.D
 Acq On : 25 Nov 2020 10:36
 Operator : JC/MD
 Sample : VN1125WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1125WBL01

Quant Time: Nov 26 03:55:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	161751	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	287226	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	297646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	134778	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.98	65	100322	53.35	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.70%	
35) Dibromofluoromethane		7.55	113	92840	54.02	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.04%	
50) Toluene-d8		10.06	98	359063	51.13	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.26%	
62) 4-Bromofluorobenzene		12.37	95	137761	54.82	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064846.D
Acq On : 25 Nov 2020 10:36
Operator : JC/MD
Sample : VN1125WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1125WBL01

Quant Time: Nov 26 03:55:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

