

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064849.D
 Acq On : 25 Nov 2020 11:52
 Operator : JC/MD
 Sample : L4798-06DL 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 45BR-20201118DL

Quant Time: Nov 26 04:01:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	158661	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	282466	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	289687	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	128619	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	97520	52.87	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	105.74%	
35) Dibromofluoromethane	7.55	113	91162	53.94	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	107.88%	
50) Toluene-d8	10.06	98	352433	51.03	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	102.06%	
62) 4-Bromofluorobenzene	12.38	95	134178	54.29	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	108.58%	

Target Compounds

					Qvalue
4) Vinyl Chloride	2.17	62	652	0.350 ug/l #	69
27) cis-1,2-Dichloroethene	6.78	96	22899	11.911 ug/l	90
44) Trichloroethene	8.80	130	67635	28.572 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064849.D
Acq On : 25 Nov 2020 11:52
Operator : JC/MD
Sample : L4798-06DL 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
45BR-20201118DL

Quant Time: Nov 26 04:01:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

