

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064850.D
 Acq On : 25 Nov 2020 12:17
 Operator : JC/MD
 Sample : L4798-12DL 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BRP2-20201118DL

Quant Time: Nov 26 04:06:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	159039	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	275793	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	287008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	127445	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	97809	52.90	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.80%	
35) Dibromofluoromethane		7.55	113	89579	54.28	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.56%	
50) Toluene-d8		10.06	98	351572	52.14	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.28%	
62) 4-Bromofluorobenzene		12.38	95	133758	55.43	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.83	85	909	0.659	ug/l	94
4) Vinyl Chloride	2.17	62	34132	18.278	ug/l	94
27) cis-1,2-Dichloroethene	6.78	96	161558	83.833	ug/l	89
44) Trichloroethene	8.80	130	42805	18.520	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064850.D
Acq On : 25 Nov 2020 12:17
Operator : JC/MD
Sample : L4798-12DL 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BRP2-20201118DL

Quant Time: Nov 26 04:06:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

