

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064852.D
 Acq On : 25 Nov 2020 13:05
 Operator : JC/MD
 Sample : L4889-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB178-TB-20201119

Quant Time: Nov 26 04:08:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	158240	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	281470	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	291532	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	131854	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	99516	54.10	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	108.20%		
35) Dibromofluoromethane	7.55	113	90487	53.73	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	107.46%		
50) Toluene-d8	10.06	98	351393	51.06	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.12%		
62) 4-Bromofluorobenzene	12.37	95	135465	55.00	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	110.00%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064852.D
Acq On : 25 Nov 2020 13:05
Operator : JC/MD
Sample : L4889-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB178-TB-20201119

Quant Time: Nov 26 04:08:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

