

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064853.D
 Acq On : 25 Nov 2020 13:29
 Operator : JC/MD
 Sample : L4889-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB178-GW-788-790

Quant Time: Nov 27 06:18:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	153337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	275329	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	285177	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	122352	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96992	54.41	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	108.82%	
35) Dibromofluoromethane	7.55	113	89427	54.28	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	108.56%	
50) Toluene-d8	10.06	98	341978	50.80	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	101.60%	
62) 4-Bromofluorobenzene	12.38	95	130694	54.25	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	108.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064853.D
Acq On : 25 Nov 2020 13:29
Operator : JC/MD
Sample : L4889-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB178-GW-788-790

Quant Time: Nov 27 06:18:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

