

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064854.D
 Acq On : 25 Nov 2020 14:17
 Operator : JC/MD
 Sample : PB133268TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133268TB

Quant Time: Nov 26 04:15:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	167966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	297914	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	322497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136144	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	102145	52.31	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	104.62%		
35) Dibromofluoromethane	7.55	113	94888	53.23	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	106.46%		
50) Toluene-d8	10.06	98	388424	53.33	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.66%		
62) 4-Bromofluorobenzene	12.38	95	149217	57.24	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	114.48%		

Target Compounds

					Qvalue
16) Acetone	3.78	43	5726	8.920 ug/l	99
20) Methylene Chloride	4.50	84	7971	3.963 ug/l	90
43) Isopropyl Acetate	8.13	43	43930	10.264 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064854.D
Acq On : 25 Nov 2020 14:17
Operator : JC/MD
Sample : PB133268TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133268TB

Quant Time: Nov 26 04:15:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

