

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064858.D
 Acq On : 25 Nov 2020 16:06
 Operator : JC/MD
 Sample : L4895-01DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-56-67DL

Quant Time: Nov 26 04:31:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	168726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	317539	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	334705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	166023	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.98	65	104584	53.32	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.64%
35) Dibromofluoromethane	7.55	113	95503	50.26	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.52%
50) Toluene-d8	10.06	98	397869	51.25	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.50%
62) 4-Bromofluorobenzene	12.37	95	169195	60.90	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	121.80%

Target Compounds

					Qvalue
16) Acetone	3.78	43	306972	476.061	ug/l 98
52) Toluene	10.12	92	102792	18.506	ug/l 100
84) 1,2,4-Trimethylbenzene	13.01	105	12214	1.259	ug/l 97
86) p-Isopropyltoluene	13.26	119	11066	1.138	ug/l 93
95) Naphthalene	15.10	128	60347	5.995	ug/l # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064858.D
Acq On : 25 Nov 2020 16:06
Operator : JC/MD
Sample : L4895-01DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-56-67DL

Quant Time: Nov 26 04:31:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

