

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112520\
 Data File : VN064854.D
 Acq On : 25 Nov 2020 14:17
 Operator : JC/MD
 Sample : PB133268TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133268TB

Quant Time: Nov 26 04:15:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	167966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	297914	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	322497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136144	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	102145	52.31	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.62%	
35) Dibromofluoromethane		7.55	113	94888	53.23	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.46%	
50) Toluene-d8		10.06	98	388424	53.33	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.66%	
62) 4-Bromofluorobenzene		12.38	95	149217	57.24	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	114.48%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	5726	8.920	ug/l	99
20) Methylene Chloride	4.50	84	7971	3.963	ug/l	90
43) Isopropyl Acetate	8.13	43	43930	10.264	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112520\
Data File : VN064854.D
Acq On : 25 Nov 2020 14:17
Operator : JC/MD
Sample : PB133268TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133268TB

Quant Time: Nov 26 04:15:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN064854.D

Acq: 25 Nov 2020 14:17

Instrument :

MSVOA_N

ClientSampleId :


PB133268TB

Tot Ion:168 Resp: 167966

Ion Ratio Lower Upper

168 100

99 52.6 43.0 64.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

80000



60000

40000

20000

0

Time-->



#16

Acetone

Concen: 8.920 ug/l

RT: 3.78 min Scan# 636

Delta R.T. 0.01 min

Lab File: VN064854.D


Acq: 25 Nov 2020 14:17

Tgt Ion: 43 Resp: 5726

Ion Ratio Lower Upper

43 100

58 31.6 25.7 38.5



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO

3.78

2500

2000


1500


1000

500

0

Time-->





#20

Methvlene Chloride

Concen: 3.963 ug/l

RT: 4.50 min Scan# 858

Delta R.T. 0.00 min

Lab File: VN064854.D

Acq: 25 Nov 2020 14:17

Instrument :

MSVOA_N

ClientSampleId :

PB133268TB

Tot Ion: 84 Resp: 7971


Ion Ratio Lower Upper

84 100

49 107.2 93.4 140.0

51 32.6 28.2 42.4

86 52.3 51.7 77.5






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: 52.309 ug/l

RT: 7.99 min Scan# 1943

Delta R.T. 0.00 min

Lab File: VN064854.D


Acq: 25 Nov 2020 14:17

Tgt Ion: 65 Resp: 102145

Ion Ratio Lower Upper



65 100


67 53.8 0.0 108.2



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.55 min Scan# 2118
Delta R.T. 0.00 min
Lab File: VN064854.D
Acq: 25 Nov 2020 14:17

Instrument :
MSVOA_N
ClientSampleId :
PB133268TB

Tot Ion:114 Resp: 297914
Ion Ratio Lower Upper
114 100
63 19.2 0.0 38.6
88 15.9 0.0 32.6




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 53.229 ug/l
RT: 7.55 min Scan# 1807
Delta R.T. 0.00 min
Lab File: VN064854.D
Acq: 25 Nov 2020 14:17

Tgt Ion:113 Resp: 94888
Ion Ratio Lower Upper
113 100
111 102.3 81.6 122.4
192 20.1 16.2 24.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 10.264 ug/l

RT: 8.13 min Scan# 1987

Delta R.T. 0.00 min

Lab File: VN064854.D

Acq: 25 Nov 2020 14:17

Instrument :

MSVOA_N

ClientSampleId :

PB133268TB


Tot Ion: 43 Resp: 43930

Ion Ratio Lower Upper

43 100

61 21.8 17.2 25.8




87 13.6 11.4 17.0



Abundance Ion 43.00 (42.70 to 43.70): VN064779.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN064779.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN064779.D (-1968) (-)



#50

Toluene-d8

Concen: 53.328 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN064854.D


Acq: 25 Nov 2020 14:17

Tgt Ion: 98 Resp: 388424

Ion Ratio Lower Upper



98 100

100 63.5 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN064779.D (-2572) (-)

Ion 100.00 (99.70 to 100.70): VN064779.D (-2572) (-)









#62
4-Bromofluorobenzene
Concen: 57.244 ug/l
RT: 12.38 min Scan# 3308
Delta R.T. 0.00 min
Lab File: VN064854.D
Acq: 25 Nov 2020 14:17




Instrument :
MSVOA_N
ClientSampleId :
PB133268TB


Tot Ion: 95 Resp: 149217
Ion Ratio Lower Upper
95 100
174 80.2 0.0 160.4
176 77.1 0.0 158.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN064854.D
Acq: 25 Nov 2020 14:17

Tgt Ion: 117 Resp: 322497
Ion Ratio Lower Upper
117 100
82 55.0 44.2 66.2
119 31.4 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3601

Delta R.T. 0.00 min

Lab File: VN064854.D


Acq: 25 Nov 2020 14:17

Instrument :

MSVOA_N

ClientSampleId :

PB133268TB




Tot Ion:152 Resp: 136144

Ion Ratio Lower Upper

152 100

115 59.2 29.7 89.1

150 157.6 0.0 350.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

