

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112522\
 Data File : VN075422.D
 Acq On : 25 Nov 2022 15:42
 Operator : JC\MD
 Sample : N5756-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT205S1-20221118

Quant Time: Nov 28 03:36:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	145082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	252407	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	216107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91502	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	39885	51.215	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.420%
35) Dibromofluoromethane	8.171	113	48633	49.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.100%
50) Toluene-d8	10.571	98	106023	47.475	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.960%
62) 4-Bromofluorobenzene	12.853	95	72205	43.804	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.600%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	9049	4.995	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112522\
Data File : VN075422.D
Acq On : 25 Nov 2022 15:42
Operator : JC\MD
Sample : N5756-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT205S1-20221118

Quant Time: Nov 28 03:36:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

