

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112522\
 Data File : VN075424.D
 Acq On : 25 Nov 2022 16:30
 Operator : JC\MD
 Sample : N5756-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW181S-20221117

Quant Time: Nov 28 03:37:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	144692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	249247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	216626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	41993	54.067	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.140%	
35) Dibromofluoromethane	8.171	113	51327	52.421	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.840%	
50) Toluene-d8	10.571	98	107609	48.796	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.600%	
62) 4-Bromofluorobenzene	12.853	95	74660	45.868	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.740%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.483	96	1458	0.844	ug/l	87
30) Chloroform	7.971	83	1602	0.549	ug/l	85
44) Trichloroethene	9.359	130	2517	1.407	ug/l	94
64) Tetrachloroethene	11.106	164	3898	2.356	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112522\
Data File : VN075424.D
Acq On : 25 Nov 2022 16:30
Operator : JC\MD
Sample : N5756-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW181S-20221117

Quant Time: Nov 28 03:37:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

