

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112522\
 Data File : VN075428.D
 Acq On : 25 Nov 2022 18:06
 Operator : JC\MD
 Sample : N5756-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW01S-20221120

Quant Time: Nov 28 03:38:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	132486	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	231084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	202961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	82303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	39774	55.928	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.860%			
35) Dibromofluoromethane	8.171	113	49981	55.059	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.120%			
50) Toluene-d8	10.571	98	104155	50.942	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.880%			
62) 4-Bromofluorobenzene	12.853	95	73174	48.488	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.980%			
Target Compounds						
29) Tetrahydrofuran	7.859	42	4455	7.908	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112522\
Data File : VN075428.D
Acq On : 25 Nov 2022 18:06
Operator : JC\MD
Sample : N5756-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01S-20221120

Quant Time: Nov 28 03:38:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

