

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112524\
 Data File : VN085022.D
 Acq On : 25 Nov 2024 10:59
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 25 15:13:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:38:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	91	0.00
2 M	Dichlorodifluoromethane	1.745	1.865	-6.9	92	0.00
3 M	Chloromethane	2.101	1.994	5.1	83	0.00
4 M	Vinyl Chloride	1.951	1.907	2.3	84	0.00
5 M	Bromomethane	0.990	1.011	-2.1	84	0.00
6 M	Chloroethane	1.281	1.395	-8.9	94	0.02
7 M	Trichlorofluoromethane	3.168	3.320	-4.8	89	0.01
8 T	Diethyl Ether	1.130	1.096	3.0	85	0.00
9	1,1,2-Trichlorotrifluoroeth	1.790	1.908	-6.6	89	0.00
10 M	1,1-Dichloroethene	1.748	1.720	1.6	85	0.00
11	Methyl Iodide	2.403	2.495	-3.8	90	0.00
12	Methyl Acetate	2.725	2.223	18.4	75	0.00
13 M	Acrolein	0.336	0.318	5.4	82	0.00
14 M	Acrylonitrile	0.892	0.834	6.5	84	0.00
15 M	Acetone	0.229	0.227	0.9	88	0.00
16 M	Carbon Disulfide	5.144	4.833	6.0	82	0.00
17	Allyl chloride	2.979	2.658	10.8	78	0.01
18 M	Methylene Chloride	2.049	2.022	1.3	83	0.00
19 M	trans-1,2-Dichloroethene	1.833	1.726	5.8	82	0.00
20 T	Diisopropyl ether	5.964	5.752	3.6	83	0.00
21 M	1,1-Dichloroethane	3.587	3.541	1.3	85	0.00
22 M	cis-1,2-Dichloroethene	2.188	2.063	5.7	84	0.00
23 M	tert-Butyl Alcohol	0.311	0.279	10.3	81	0.00
24 M	Methyl tert-Butyl Ether	5.654	5.397	4.5	84	0.00
25 M	Chloroform	3.675	3.651	0.7	85	0.00
26	Cyclohexane	2.740	2.742	-0.1	88	0.00
27 s	1,2-Dichloroethane-d4	2.412	2.453	-1.7	90	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
29	1,1-Dichloropropene	0.452	0.450	0.4	85	0.00
30 M	2-Butanone	0.208	0.198	4.8	84	0.00
31	2,2-Dichloropropane	0.552	0.591	-7.1	89	0.00
32 M	1,1,1-Trichloroethane	0.614	0.627	-2.1	86	0.00
33 M	Carbon Tetrachloride	0.524	0.544	-3.8	89	0.00
34 M	Benzene	1.494	1.470	1.6	85	0.00
35	Methacrylonitrile	0.247	0.239	3.2	86	0.00
36 M	1,2-Dichloroethane	0.497	0.504	-1.4	85	0.00
37 M	Trichloroethene	0.337	0.335	0.6	84	0.00
38	Methylcyclohexane	0.486	0.476	2.1	88	0.00
39 M	1,2-Dichloropropane	0.360	0.358	0.6	85	0.00
40	Dibromomethane	0.252	0.245	2.8	82	0.00
41 M	Bromodichloromethane	0.539	0.534	0.9	86	0.00
42 M	Vinyl Acetate	0.779	0.745	4.4	84	0.00
43	Ethyl Acetate	0.438	0.409	6.6	84	0.00
44	Isopropyl Acetate	0.887	0.707	20.3	74	0.00
45 T	1,4-Dioxane	0.006	0.006	0.0	84	0.00
46	Methyl methacrylate	0.356	0.312	12.4	79	0.00
47	n-amyl Acetate	0.662	0.562	15.1	74	0.00
48 M	t-1,3-Dichloropropene	0.532	0.499	6.2	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112524\
 Data File : VN085022.D
 Acq On : 25 Nov 2024 10:59
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 25 15:13:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:38:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.571	0.561	1.8	85	0.00
50 M	1,1,2-Trichloroethane	0.337	0.339	-0.6	84	0.00
51	Ethyl methacrylate	0.520	0.473	9.0	83	0.00
52	1,3-Dichloropropane	0.586	0.572	2.4	83	0.00
53 M	Dibromochloromethane	0.395	0.392	0.8	86	0.00
54 M	1,2-Dibromoethane	0.341	0.337	1.2	83	0.00
55 M	2-Chloroethyl vinyl ether	0.255	0.203	20.4	67	0.00
56 M	Bromoform	0.265	0.258	2.6	87	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
58 M	4-Methyl-2-Pentanone	0.473	0.432	8.7	81	0.00
59 M	2-Hexanone	0.347	0.321	7.5	83	0.00
60 S	4-Bromofluorobenzene	0.517	0.534	-3.3	96	0.00
61 M	Tetrachloroethene	0.314	0.316	-0.6	87	0.00
62 M	Toluene	1.709	1.640	4.0	84	0.00
63 S	Toluene-d8	1.420	1.409	0.8	89	0.00
64 M	Chlorobenzene	1.032	0.997	3.4	85	0.00
65	1,1,1,2-Tetrachloroethane	0.379	0.383	-1.1	88	0.00
66 M	Ethyl Benzene	1.796	1.679	6.5	84	0.00
67 M	m/p-Xylenes	0.687	0.664	3.3	85	0.00
68 M	o-Xylene	0.663	0.616	7.1	81	0.00
69 M	Styrene	1.112	1.046	5.9	84	0.00
70	Isopropylbenzene	1.647	1.585	3.8	87	0.00
71 M	1,1,2,2-Tetrachloroethane	0.534	0.511	4.3	82	0.00
72	1,2,3-Trichloropropane	0.456	0.472	-3.5	82	0.00
73	Bromobenzene	0.425	0.421	0.9	87	0.00
74	n-propylbenzene	1.926	1.861	3.4	87	0.00
75	2-Chlorotoluene	1.255	1.229	2.1	87	0.00
76	1,3,5-Trimethylbenzene	1.408	1.376	2.3	87	0.00
77	t-1,4-Dichloro-2-butene	0.187	0.171	8.6	82	0.00
78	4-Chlorotoluene	1.268	1.190	6.2	84	0.00
79	tert-butylbenzene	1.164	1.130	2.9	90	0.00
80	1,2,4-Trimethylbenzene	1.420	1.411	0.6	89	0.00
81	sec-Butylbenzene	1.604	1.574	1.9	90	0.00
82	p-Isopropyltoluene	1.348	1.331	1.3	91	0.00
83 M	1,3-Dichlorobenzene	0.786	0.751	4.5	84	0.00
84 M	1,4-Dichlorobenzene	0.776	0.721	7.1	84	0.00
85	n-Butylbenzene	1.132	1.069	5.6	91	0.00
86 T	Hexachloroethane	0.276	0.272	1.4	87	0.00
87 M	1,2-Dichlorobenzene	0.774	0.734	5.2	82	0.00
88	1,2-Dibromo-3-Chloropropane	0.103	0.097	5.8	82	0.00
89	1,2,4-Trichlorobenzene	0.377	0.338	10.3	81	0.00
90	Hexachlorobutadiene	0.165	0.183	-10.9	99	0.00
91 M	Naphthalene	1.204	0.966	19.8	78	0.00
92	1,2,3-Trichlorobenzene	0.377	0.338	10.3	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0