

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112618\
 Data File : VN052461.D
 Acq On : 26 Nov 2018 16:17
 Operator : MD\SY
 Sample : VN1126WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1126WBL01

Quant Time: Nov 27 01:05:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	174206	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	1004909	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	825757	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	358395	30.19	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.63%
60) 4-Bromofluorobenzene	12.40	95	335800	26.60	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	88.67%
63) Toluene-d8	10.09	98	1185080	30.95	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	103.17%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112618\
Data File : VN052461.D
Acq On : 26 Nov 2018 16:17
Operator : MD\SY
Sample : VN1126WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1126WBL01

Quant Time: Nov 27 01:05:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 27 00:48:33 2018
Response via : Initial Calibration

