

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112618\
 Data File : VN052460.D
 Acq On : 26 Nov 2018 15:42
 Operator : MD\SY
 Sample : VSTDICV020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN112618

Quant Time: Nov 27 01:02:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	99	0.00
2 M	Dichlorodifluoromethane	1.709	1.752	-2.5	99	0.00
3 M	Chloromethane	2.266	2.301	-1.5	100	0.00
4 M	Vinyl Chloride	2.287	2.340	-2.3	101	0.00
5 M	Bromomethane	1.304	1.480	-13.5	114	0.00
6 M	Chloroethane	1.360	1.414	-4.0	102	0.00
7 M	Trichlorofluoromethane	2.823	2.932	-3.9	103	0.00
8 T	Diethyl Ether	1.042	1.070	-2.7	105	0.00
9	1,1,2-Trichlorotrifluoroeth	1.789	1.853	-3.6	100	0.00
10 M	1,1-Dichloroethene	1.682	1.748	-3.9	104	0.00
11	Methyl Iodide	2.336	2.071	11.3	91	0.00
12	Methyl Acetate	1.429	1.412	1.2	97	0.00
13 M	Acrolein	0.168	0.151	10.1	97	0.00
14 M	Acrylonitrile	0.609	0.595	2.3	95	0.00
15 M	Acetone	0.158	0.144	8.9	83	0.00
16 M	Carbon Disulfide	5.326	5.484	-3.0	103	0.00
17	Allyl chloride	2.934	3.029	-3.2	102	0.00
18 M	Methylene Chloride	1.948	2.003	-2.8	103	0.00
19 M	trans-1,2-Dichloroethene	1.810	1.881	-3.9	104	0.00
20 T	Diisopropyl ether	6.166	6.274	-1.8	100	0.00
21 M	1,1-Dichloroethane	3.514	3.627	-3.2	103	0.00
22 M	cis-1,2-Dichloroethene	2.075	2.094	-0.9	100	0.00
23 M	tert-Butyl Alcohol	0.091	0.071	22.0	76	0.00
24 M	Methyl tert-Butyl Ether	4.413	4.470	-1.3	101	0.00
25 M	Chloroform	3.390	3.481	-2.7	102	0.00
26	Cyclohexane	3.314	3.449	-4.1	103	0.00
27 s	1,2-Dichloroethane-d4	2.044	2.036	0.4	99	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
29	1,1-Dichloropropene	0.458	0.468	-2.2	103	0.00
30 M	2-Butanone	0.123	0.113	8.1	91	0.00
31	2,2-Dichloropropane	0.422	0.426	-0.9	105	0.00
32 M	1,1,1-Trichloroethane	0.477	0.472	1.0	102	0.00
33 M	Carbon Tetrachloride	0.417	0.421	-1.0	103	0.00
34 M	Benzene	1.377	1.388	-0.8	102	0.00
35	Methacrylonitrile	0.152	0.139	8.6	92	0.00
36 M	1,2-Dichloroethane	0.417	0.411	1.4	100	0.00
37 M	Trichloroethene	0.349	0.350	-0.3	103	0.00
38	Methylcyclohexane	0.567	0.589	-3.9	104	0.00
39 M	1,2-Dichloropropane	0.369	0.366	0.8	101	0.00
40	Dibromomethane	0.200	0.201	-0.5	102	0.00
41 M	Bromodichloromethane	0.441	0.432	2.0	101	0.00
42 M	Vinyl Acetate	0.663	0.651	1.8	101	0.00
43	Ethyl Acetate	0.261	0.248	5.0	94	0.00
44	Isopropyl Acetate	0.457	0.444	2.8	100	0.00
45 T	1,4-Dioxane	0.003	0.002#	33.3#	74	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112618\
 Data File : VN052460.D
 Acq On : 26 Nov 2018 15:42
 Operator : MD\SY
 Sample : VSTDICV020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleID :
 ICVVN112618

Quant Time: Nov 27 01:02:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.237	0.229	3.4	99	0.00
47	n-amyl Acetate	0.379	0.356	6.1	101	0.00
48 M	t-1,3-Dichloropropene	0.427	0.418	2.1	105	0.00
49 T	cis-1,3-Dichloropropene	0.521	0.514	1.3	102	0.00
50 M	1,1,2-Trichloroethane	0.283	0.277	2.1	98	0.00
51	Ethyl methacrylate	0.355	0.345	2.8	102	0.00
52	1,3-Dichloropropane	0.503	0.487	3.2	98	0.00
53 M	Dibromochloromethane	0.311	0.302	2.9	99	0.00
54 M	1,2-Dibromoethane	0.276	0.275	0.4	102	0.00
55 M	2-Chloroethyl vinyl ether	0.208	0.200	3.8	96	0.00
56 M	Bromoform	0.204	0.194	4.9	99	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
58 M	4-Methyl-2-Pentanone	0.291	0.280	3.8	94	0.00
59 M	2-Hexanone	0.199	0.189	5.0	92	0.00
60 S	4-Bromofluorobenzene	0.459	0.458	0.2	101	0.00
61 M	Tetrachloroethene	0.369	0.382	-3.5	101	0.00
62 M	Toluene	1.654	1.675	-1.3	101	0.00
63 S	Toluene-d8	1.391	1.408	-1.2	101	0.00
64 M	Chlorobenzene	1.009	1.035	-2.6	103	0.00
65	1,1,1,2-Tetrachloroethane	0.350	0.352	-0.6	103	0.00
66 M	Ethyl Benzene	1.782	1.811	-1.6	102	0.00
67 M	m/p-Xylenes	0.664	0.688	-3.6	103	0.00
68 M	o-Xylene	0.645	0.657	-1.9	102	0.00
69 M	Styrene	1.040	1.060	-1.9	102	0.00
70	Isopropylbenzene	1.715	1.778	-3.7	104	0.00
71 M	1,1,2,2-Tetrachloroethane	0.373	0.365	2.1	96	0.00
72	1,2,3-Trichloropropane	0.320	0.290	9.4	95	0.00
73	Bromobenzene	0.415	0.425	-2.4	104	0.00
74	n-propylbenzene	1.998	2.069	-3.6	103	0.00
75	2-Chlorotoluene	1.160	1.203	-3.7	103	0.00
76	1,3,5-Trimethylbenzene	1.385	1.461	-5.5	104	0.00
77	t-1,4-Dichloro-2-butene	0.101	0.095	5.9	100	0.00
78	4-Chlorotoluene	1.158	1.195	-3.2	103	0.00
79	tert-butylbenzene	1.206	1.271	-5.4	104	0.00
80	1,2,4-Trimethylbenzene	1.386	1.453	-4.8	103	0.00
81	sec-Butylbenzene	1.680	1.762	-4.9	104	0.00
82	p-Isopropyltoluene	1.431	1.505	-5.2	105	0.00
83 M	1,3-Dichlorobenzene	0.729	0.764	-4.8	106	0.00
84 M	1,4-Dichlorobenzene	0.705	0.734	-4.1	105	0.00
85	n-Butylbenzene	1.229	1.282	-4.3	105	0.00
86 T	Hexachloroethane	0.242	0.246	-1.7	103	0.00
87 M	1,2-Dichlorobenzene	0.694	0.720	-3.7	103	0.00
88	1,2-Dibromo-3-Chloropropane	0.052	0.051	1.9	100	0.00
89	1,2,4-Trichlorobenzene	0.342	0.355	-3.8	108	0.00
90	Hexachlorobutadiene	0.224	0.267	-19.2	112	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112618\
Data File : VN052460.D
Acq On : 26 Nov 2018 15:42
Operator : MD\SY
Sample : VSTDICV020
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
ICVVN112618

Quant Time: Nov 27 01:02:09 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 27 00:48:33 2018
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	0.691	0.714	-3.3	108	0.00
92	1,2,3-Trichlorobenzene	0.315	0.332	-5.4	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0