

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112621\
 Data File : VN069650.D
 Acq On : 26 Nov 2021 11:02
 Operator : JC/MD
 Sample : VN1126WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1126WBL01

Quant Time: Nov 29 00:15:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Nov 10 14:36:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	192993	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	1110713	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	923400	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.437	65	543906	32.496	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	108.333%
60) 4-Bromofluorobenzene	12.734	95	364227	24.648	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.167%
63) Toluene-d8	10.443	98	1332084	31.243	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.133%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112621\
Data File : VN069650.D
Acq On : 26 Nov 2021 11:02
Operator : JC/MD
Sample : VN1126WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1126WBL01

Quant Time: Nov 29 00:15:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111021W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Nov 10 14:36:17 2021
Response via : Initial Calibration

