

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112624\
 Data File : VN085043.D
 Acq On : 26 Nov 2024 15:23
 Operator : JC\MD
 Sample : PB165249TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB165249TB

Quant Time: Nov 27 05:15:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	146335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262989	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	243699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	108433	52.020	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	104.040%	
35) Dibromofluoromethane	8.165	113	87725	49.464	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.920%	
50) Toluene-d8	10.565	98	314488	49.480	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.960%	
62) 4-Bromofluorobenzene	12.847	95	116617	49.063	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	98.120%	
Target Compounds						
16) Acetone	4.430	43	30235	49.805	ug/l	98
20) Methylene Chloride	5.277	84	2979	1.608	ug/l #	74
43) Isopropyl Acetate	8.688	43	137728	30.362	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112624\
Data File : VN085043.D
Acq On : 26 Nov 2024 15:23
Operator : JC\MD
Sample : PB165249TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB165249TB

Quant Time: Nov 27 05:15:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

