

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112624\
 Data File : VN085045.D
 Acq On : 26 Nov 2024 16:11
 Operator : JC\MD
 Sample : P4985-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-740-WC

Quant Time: Nov 27 05:16:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	160713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	109090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116674	50.966	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.940%	
35) Dibromofluoromethane	8.165	113	95094	48.466	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.940%	
50) Toluene-d8	10.565	98	338687	48.166	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	96.340%	
62) 4-Bromofluorobenzene	12.847	95	127670	48.551	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	97.100%	
Target Compounds						
16) Acetone	4.424	43	57325	85.981	ug/l	99
18) Methyl Acetate	5.024	43	4945	1.201	ug/l	93
20) Methylene Chloride	5.277	84	2865	1.408	ug/l #	76
25) 2-Butanone	7.494	43	5875	5.390	ug/l	90
43) Isopropyl Acetate	8.688	43	231141	46.057	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112624\
Data File : VN085045.D
Acq On : 26 Nov 2024 16:11
Operator : JC\MD
Sample : P4985-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-740-WC

Quant Time: Nov 27 05:16:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

