

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112624\
 Data File : VN085049.D
 Acq On : 26 Nov 2024 17:48
 Operator : JC\MD
 Sample : P5006-18 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-WTS-02

Quant Time: Nov 27 05:19:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	144182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	264529	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	220546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	93934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	108530	52.844	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.680%			
35) Dibromofluoromethane	8.165	113	86136	48.285	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.580%			
50) Toluene-d8	10.565	98	294207	46.020	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.040%			
62) 4-Bromofluorobenzene	12.847	95	105504	44.129	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 88.260%			
Target Compounds						
					Qvalue	
18) Methyl Acetate	5.024	43	4846	1.420	ug/l #	86
68) m/p-Xylenes	12.076	106	896	0.290	ug/l	94
70) Styrene	12.417	104	1285	0.263	ug/l #	80
77) Bromobenzene	12.982	156	733	0.411	ug/l	86
78) n-propylbenzene	13.035	91	2443	0.320	ug/l	89
82) 4-Chlorotoluene	13.217	91	2200	0.414	ug/l	96
89) n-Butylbenzene	14.053	91	2128	0.429	ug/l	96
91) 1,2-Dichlorobenzene	14.106	146	1586	0.471	ug/l	95
93) 1,2,4-Trichlorobenzene	15.388	180	1082	0.591	ug/l	75
95) Naphthalene	15.641	128	3912	0.750	ug/l #	86
96) 1,2,3-Trichlorobenzene	15.835	180	1321	0.756	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112624\
Data File : VN085049.D
Acq On : 26 Nov 2024 17:48
Operator : JC\MD
Sample : P5006-18 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-WTS-02

Quant Time: Nov 27 05:19:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

