

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112625\
 Data File : VN088365.D
 Acq On : 26 Nov 2025 10:36
 Operator : JC\MD
 Sample : VN1126WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1126WBL01

Quant Time: Nov 27 00:10:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111325W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 14 00:49:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	48647	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	246777	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	220989	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	127683	28.734	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	95.767%
60) 4-Bromofluorobenzene	12.829	95	92288	24.704	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.333%
63) Toluene-d8	10.547	98	304303	29.085	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.967%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112625\
Data File : VN088365.D
Acq On : 26 Nov 2025 10:36
Operator : JC\MD
Sample : VN1126WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1126WBL01

Quant Time: Nov 27 00:10:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111325W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 14 00:49:21 2025
Response via : Initial Calibration

