

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
 Data File : VN052481.D
 Acq On : 27 Nov 2018 16:20
 Operator : MD\SY
 Sample : J6063-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 28 06:44:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1329958	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2034857	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1727144	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	627386	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	678825	46.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.76%	
35) Dibromofluoromethane	7.59	113	611054	48.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.42%	
50) Toluene-d8	10.09	98	2450765	50.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.40	95	759709	46.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.60%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
Data File : VN052481.D
Acq On : 27 Nov 2018 16:20
Operator : MD\SY
Sample : J6063-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 28 06:44:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

