

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
 Data File : VN052484.D
 Acq On : 27 Nov 2018 17:40
 Operator : MD\SY
 Sample : PB115074ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115074ZHE#01

Quant Time: Nov 28 06:49:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1324696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2035845	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1766756	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	654272	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	688240	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
35) Dibromofluoromethane	7.59	113	614631	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
50) Toluene-d8	10.09	98	2483804	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
62) 4-Bromofluorobenzene	12.40	95	781006	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
Target Compounds						
16) Acetone	3.82	43	31561	8.30	ug/l	90
20) Methylene Chloride	4.55	84	29106	1.96	ug/l	92
43) Isopropyl Acetate	8.17	43	44038	2.22	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
Data File : VN052484.D
Acq On : 27 Nov 2018 17:40
Operator : MD\SY
Sample : PB115074ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115074ZHE#01

Quant Time: Nov 28 06:49:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

