

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
 Data File : VN052485.D
 Acq On : 27 Nov 2018 18:06
 Operator : MD\SY
 Sample : PB115074ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115074ZHE#02

Quant Time: Nov 28 06:50:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1276217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1968377	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1676260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	571649	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	666102	47.94	ug/l	0.00
Spiked Amount			Recovery	=	95.88%	
35) Dibromofluoromethane	7.59	113	599294	49.39	ug/l	0.00
Spiked Amount			Recovery	=	98.78%	
50) Toluene-d8	10.09	98	2406713	50.95	ug/l	0.00
Spiked Amount			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.40	95	707299	44.56	ug/l	0.00
Spiked Amount			Recovery	=	89.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	29481	8.04	ug/l	90
20) Methylene Chloride	4.55	84	28953	2.02	ug/l	93
43) Isopropyl Acetate	8.17	43	43808	2.28	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
Data File : VN052485.D
Acq On : 27 Nov 2018 18:06
Operator : MD\SY
Sample : PB115074ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115074ZHE#02

Quant Time: Nov 28 06:50:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

