

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
 Data File : VN052500.D
 Acq On : 28 Nov 2018 1:38
 Operator : MD\SY
 Sample : J6086-01 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-1

Quant Time: Nov 28 09:02:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	172217	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	987979	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	816339	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	357304	30.45	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	101.50%	
60) 4-Bromofluorobenzene	12.40	95	329112	26.37	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	87.90%	
63) Toluene-d8	10.09	98	1154538	30.50	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	101.67%	

Target Compounds

						Qvalue
25) Chloroform	7.37	83	3983560	204.69	ug/l	99
41) Bromodichloromethane	9.40	83	397202	27.36	ug/l	97
53) Dibromochloromethane	10.90	129	17341	1.69	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112718\
Data File : VN052500.D
Acq On : 28 Nov 2018 1:38
Operator : MD\SY
Sample : J6086-01 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
14B-1

Quant Time: Nov 28 09:02:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N112618W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 27 00:48:33 2018
Response via : Initial Calibration

