

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112723\
 Data File : VN080267.D
 Acq On : 27 Nov 2023 11:17
 Operator : JC\MD
 Sample : VN1127MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1127MBL01

Quant Time: Nov 28 02:06:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	139433	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	250517	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	266497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115418	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	126717	55.040	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.080%	
35) Dibromofluoromethane	8.165	113	98470	53.271	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.540%	
50) Toluene-d8	10.570	98	336209	50.456	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.920%	
62) 4-Bromofluorobenzene	12.853	95	127016	50.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.940%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112723\
Data File : VN080267.D
Acq On : 27 Nov 2023 11:17
Operator : JC\MD
Sample : VN1127MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1127MBL01

Quant Time: Nov 28 02:06:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

