

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112723\
 Data File : VN080283.D
 Acq On : 27 Nov 2023 18:23
 Operator : JC\MD
 Sample : 05546-02 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1146

Quant Time: Nov 28 02:10:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	285327	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306909	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	144169	55.450	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.900%
35) Dibromofluoromethane	8.171	113	113051	53.697	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.400%
50) Toluene-d8	10.571	98	379107	49.953	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.900%
62) 4-Bromofluorobenzene	12.853	95	152167	53.083	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.160%
Target Compounds						
						Qvalue
52) Toluene	10.630	92	2272	0.419	ug/l	91
68) m/p-Xylenes	12.071	106	2161	0.479	ug/l	92
69) o-Xylene	12.394	106	1042	0.245	ug/l	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112723\
Data File : VN080283.D
Acq On : 27 Nov 2023 18:23
Operator : JC\MD
Sample : 05546-02 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1146

Quant Time: Nov 28 02:10:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

