

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112723\
 Data File : VN080285.D
 Acq On : 27 Nov 2023 19:10
 Operator : JC\MD
 Sample : 05551-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB11272023

Quant Time: Nov 28 02:11:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	168235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	306991	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	327534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	140110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150279	54.100	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.200%	
35) Dibromofluoromethane	8.165	113	118521	52.323	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.640%	
50) Toluene-d8	10.571	98	407809	49.943	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.880%	
62) 4-Bromofluorobenzene	12.853	95	157916	51.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.400%	
Target Compounds						
					Qvalue	
3) Chloromethane	2.359	50	1642	0.872	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112723\
Data File : VN080285.D
Acq On : 27 Nov 2023 19:10
Operator : JC\MD
Sample : 05551-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB11272023

Quant Time: Nov 28 02:11:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

