

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112724\
 Data File : VN085057.D
 Acq On : 27 Nov 2024 14:20
 Operator : JC\MD
 Sample : P5018-02 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2

Quant Time: Nov 28 06:57:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:38:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.806	128	21362	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.094	114	109546	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	94649	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.576	65	51084	29.745	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.167%
60) 4-Bromofluorobenzene	12.847	95	38956	23.904	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	79.667%
63) Toluene-d8	10.565	98	125867	28.091	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	93.633%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112724\
Data File : VN085057.D
Acq On : 27 Nov 2024 14:20
Operator : JC\MD
Sample : P5018-02 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2

Quant Time: Nov 28 06:57:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 01 02:38:04 2024
Response via : Initial Calibration

