

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112818\
 Data File : VN052509.D
 Acq On : 28 Nov 2018 10:29
 Operator : MD\SY
 Sample : J6084-01 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 MS-SFVP8283LM

Quant Time: Nov 29 00:17:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1260781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1933588	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1649258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	647178	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	659364	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
35) Dibromofluoromethane	7.59	113	590945	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
50) Toluene-d8	10.09	98	2328705	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.40	95	723599	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
Target Compounds						
68) m/p-Xylenes	11.62	106	9512	0.429	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	23099	0.569	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112818\
Data File : VN052509.D
Acq On : 28 Nov 2018 10:29
Operator : MD\SY
Sample : J6084-01 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
MS-SFVP8283LM

Quant Time: Nov 29 00:17:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

