

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112818\
 Data File : VN052518.D
 Acq On : 28 Nov 2018 14:35
 Operator : MD\SY
 Sample : J6084-01 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 MS-SFVP8283LM

Quant Time: Nov 29 12:30:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1560041	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2359339	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2072950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	940473	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	768634	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
35) Dibromofluoromethane	7.59	113	702941	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	10.09	98	2859868	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
62) 4-Bromofluorobenzene	12.40	95	963165	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.04	105	155956	2.644	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112818\
Data File : VN052518.D
Acq On : 28 Nov 2018 14:35
Operator : MD\SY
Sample : J6084-01 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
MS-SFVP8283LM

Quant Time: Nov 29 12:30:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

