

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112818\
 Data File : VN052528.D
 Acq On : 28 Nov 2018 19:01
 Operator : MD\SY
 Sample : J6050-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 DTW-02

Quant Time: Nov 29 01:39:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 28 01:11:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1382010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2078092	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1768705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	674040	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	675839	44.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.82%	
35) Dibromofluoromethane	7.59	113	622351	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	10.09	98	2496146	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	12.40	95	775861	46.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.60%	
Target Compounds						
16) Acetone	3.83	43	6460	1.628	ug/l #	86
52) Toluene	10.16	92	39963	1.150	ug/l	96
68) m/p-Xylenes	11.62	106	29314	1.233	ug/l	99
69) o-Xylene	11.95	106	10621	0.460	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	23901	0.565	ug/l	99
95) Naphthalene	15.14	128	76507	6.237	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112818\
Data File : VN052528.D
Acq On : 28 Nov 2018 19:01
Operator : MD\SY
Sample : J6050-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
DTW-02

Quant Time: Nov 29 01:39:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 28 01:11:38 2018
Response via : Initial Calibration

