

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112818\
 Data File : VN052533.D
 Acq On : 28 Nov 2018 21:41
 Operator : MD\SY
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 29 02:39:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	110	0.00
2 M	Dichlorodifluoromethane	1.709	1.544	9.7	96	0.00
3 M	Chloromethane	2.266	2.023	10.7	97	0.00
4 M	Vinyl Chloride	2.287	2.137	6.6	102	0.00
5 M	Bromomethane	1.304	1.224	6.1	104	0.00
6 M	Chloroethane	1.360	1.321	2.9	105	0.00
7 M	Trichlorofluoromethane	2.823	2.707	4.1	105	0.00
8 T	Diethyl Ether	1.042	1.019	2.2	110	0.00
9	1,1,2-Trichlorotrifluoroeth	1.789	1.655	7.5	98	0.00
10 M	1,1-Dichloroethene	1.682	1.648	2.0	108	0.00
11	Methyl Iodide	2.336	2.076	11.1	101	0.00
12	Methyl Acetate	1.429	1.334	6.6	102	0.00
13 M	Acrolein	0.168	0.144	14.3	102	0.00
14 M	Acrylonitrile	0.609	0.565	7.2	99	0.00
15 M	Acetone	0.158	0.127	19.6	81	0.00
16 M	Carbon Disulfide	5.326	4.734	11.1	98	0.00
17	Allyl chloride	2.934	2.668	9.1	99	0.00
18 M	Methylene Chloride	1.948	1.879	3.5	107	0.00
19 M	trans-1,2-Dichloroethene	1.810	1.742	3.8	107	0.00
20 T	Diisopropyl ether	6.166	5.774	6.4	102	0.00
21 M	1,1-Dichloroethane	3.514	3.354	4.6	105	0.00
22 M	cis-1,2-Dichloroethene	2.075	2.006	3.3	106	0.00
23 M	tert-Butyl Alcohol	0.091	0.097	-6.6	114	0.00
24 M	Methyl tert-Butyl Ether	4.413	4.513	-2.3	112	0.00
25 M	Chloroform	3.390	3.269	3.6	106	0.00
26	Cyclohexane	3.314	3.064	7.5	101	0.00
27 s	1,2-Dichloroethane-d4	2.044	1.980	3.1	106	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	0.00
29	1,1-Dichloropropene	0.458	0.427	6.8	104	0.00
30 M	2-Butanone	0.123	0.108	12.2	95	0.00
31	2,2-Dichloropropane	0.422	0.372	11.8	101	0.00
32 M	1,1,1-Trichloroethane	0.477	0.454	4.8	108	0.00
33 M	Carbon Tetrachloride	0.417	0.390	6.5	105	0.00
34 M	Benzene	1.377	1.309	4.9	106	0.00
35	Methacrylonitrile	0.152	0.149	2.0	108	0.00
36 M	1,2-Dichloroethane	0.417	0.396	5.0	106	0.00
37 M	Trichloroethene	0.349	0.340	2.6	109	0.00
38	Methylcyclohexane	0.567	0.514	9.3	99	0.00
39 M	1,2-Dichloropropane	0.369	0.341	7.6	103	0.00
40	Dibromomethane	0.200	0.191	4.5	106	0.00
41 M	Bromodichloromethane	0.441	0.412	6.6	106	0.00
42 M	Vinyl Acetate	0.663	0.606	8.6	103	0.00
43	Ethyl Acetate	0.261	0.242	7.3	100	0.00
44	Isopropyl Acetate	0.457	0.443	3.1	110	0.00
45 T	1,4-Dioxane	0.003	0.003#	0.0	113	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112818\
 Data File : VN052533.D
 Acq On : 28 Nov 2018 21:41
 Operator : MD\SY
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 29 02:39:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.237	0.226	4.6	108	0.00
47	n-amyl Acetate	0.379	0.342	9.8	106	0.00
48 M	t-1,3-Dichloropropene	0.427	0.389	8.9	107	0.00
49 T	cis-1,3-Dichloropropene	0.521	0.476	8.6	104	0.00
50 M	1,1,2-Trichloroethane	0.283	0.275	2.8	106	0.00
51	Ethyl methacrylate	0.355	0.344	3.1	112	0.00
52	1,3-Dichloropropane	0.503	0.478	5.0	105	0.00
53 M	Dibromochloromethane	0.311	0.290	6.8	105	0.00
54 M	1,2-Dibromoethane	0.276	0.268	2.9	109	0.00
55 M	2-Chloroethyl vinyl ether	0.208	0.196	5.8	104	0.00
56 M	Bromoform	0.204	0.187	8.3	105	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
58 M	4-Methyl-2-Pentanone	0.291	0.284	2.4	104	0.00
59 M	2-Hexanone	0.199	0.183	8.0	97	0.00
60 S	4-Bromofluorobenzene	0.459	0.457	0.4	110	0.00
61 M	Tetrachloroethene	0.369	0.372	-0.8	108	0.00
62 M	Toluene	1.654	1.640	0.8	108	0.00
63 S	Toluene-d8	1.391	1.390	0.1	109	0.00
64 M	Chlorobenzene	1.009	0.988	2.1	108	0.00
65	1,1,1,2-Tetrachloroethane	0.350	0.344	1.7	109	0.00
66 M	Ethyl Benzene	1.782	1.721	3.4	106	0.00
67 M	m/p-Xylenes	0.664	0.655	1.4	107	0.00
68 M	o-Xylene	0.645	0.640	0.8	109	0.00
69 M	Styrene	1.040	1.010	2.9	106	0.00
70	Isopropylbenzene	1.715	1.697	1.0	108	0.00
71 M	1,1,2,2-Tetrachloroethane	0.373	0.367	1.6	105	0.00
72	1,2,3-Trichloropropane	0.320	0.308	3.8	111	0.00
73	Bromobenzene	0.415	0.411	1.0	110	0.00
74	n-propylbenzene	1.998	1.885	5.7	102	0.00
75	2-Chlorotoluene	1.160	1.126	2.9	105	0.00
76	1,3,5-Trimethylbenzene	1.385	1.357	2.0	106	0.00
77	t-1,4-Dichloro-2-butene	0.101	0.086	14.9	99	0.00
78	4-Chlorotoluene	1.158	1.093	5.6	103	0.00
79	tert-butylbenzene	1.206	1.203	0.2	108	0.00
80	1,2,4-Trimethylbenzene	1.386	1.370	1.2	106	0.00
81	sec-Butylbenzene	1.680	1.602	4.6	103	0.00
82	p-Isopropyltoluene	1.431	1.361	4.9	104	0.00
83 M	1,3-Dichlorobenzene	0.729	0.705	3.3	106	0.00
84 M	1,4-Dichlorobenzene	0.705	0.684	3.0	107	0.00
85	n-Butylbenzene	1.229	1.096	10.8	98	0.00
86 T	Hexachloroethane	0.242	0.218	9.9	100	0.00
87 M	1,2-Dichlorobenzene	0.694	0.686	1.2	107	0.00
88	1,2-Dibromo-3-Chloropropane	0.052	0.051	1.9	110	0.00
89	1,2,4-Trichlorobenzene	0.342	0.324	5.3	108	0.00
90	Hexachlorobutadiene	0.224	0.214	4.5	98	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112818\
Data File : VN052533.D
Acq On : 28 Nov 2018 21:41
Operator : MD\SY
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 28

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Nov 29 02:39:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 27 00:48:33 2018
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	0.691	0.680	1.6	112	0.00
92	1,2,3-Trichlorobenzene	0.315	0.305	3.2	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0