

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
 Data File : VN075442.D
 Acq On : 28 Nov 2022 12:31
 Operator : JC\MD
 Sample : N5756-02RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW8-MW01S-20221119RE

Quant Time: Nov 29 05:31:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	115597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	194605	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	166944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	67579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	31494	50.755	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.520%
35) Dibromofluoromethane	8.177	113	37920	49.603	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.200%
50) Toluene-d8	10.570	98	76217	44.266	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.540%
62) 4-Bromofluorobenzene	12.853	95	50617	39.828	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.660%#
Target Compounds						
52) Toluene	10.623	92	1498	0.455	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
Data File : VN075442.D
Acq On : 28 Nov 2022 12:31
Operator : JC\MD
Sample : N5756-02RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW8-MW01S-20221119RE

Quant Time: Nov 29 05:31:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

