

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
 Data File : VN075443.D
 Acq On : 28 Nov 2022 12:55
 Operator : JC\MD
 Sample : N5756-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW8-MW01D2-20221120

Quant Time: Nov 29 05:31:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	137886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	232389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	202805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80432	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	40333	54.493	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.980%
35) Dibromofluoromethane	8.171	113	49087	53.770	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.540%
50) Toluene-d8	10.571	98	100824	49.036	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.853	95	69332	45.684	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
Data File : VN075443.D
Acq On : 28 Nov 2022 12:55
Operator : JC\MD
Sample : N5756-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW8-MW01D2-20221120

Quant Time: Nov 29 05:31:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

