

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
 Data File : VN075445.D
 Acq On : 28 Nov 2022 13:47
 Operator : JC\MD
 Sample : N5757-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01-20221121

Quant Time: Nov 29 05:31:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	136665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	238384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	205353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80801	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	42191	57.512	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.020%
35) Dibromofluoromethane	8.171	113	50194	53.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.200%
50) Toluene-d8	10.571	98	106204	50.354	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.700%
62) 4-Bromofluorobenzene	12.853	95	79982	51.377	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.760%
Target Compounds						
16) Acetone	4.459	43	3408	4.981	ug/l	92
25) 2-Butanone	7.489	43	3325	3.480	ug/l	92
52) Toluene	10.624	92	1280	0.318	ug/l	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
Data File : VN075445.D
Acq On : 28 Nov 2022 13:47
Operator : JC\MD
Sample : N5757-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01-20221121

Quant Time: Nov 29 05:31:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

