

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
 Data File : VN075446.D
 Acq On : 28 Nov 2022 14:11
 Operator : JC\MD
 Sample : N5757-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE137D7-20221121

Quant Time: Nov 29 05:31:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.229	168	123409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	228373	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	197902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	75804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	39364	59.423	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 118.840%	
35) Dibromofluoromethane	8.171	113	47285	52.707	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.420%	
50) Toluene-d8	10.570	98	101278	50.123	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.240%	
62) 4-Bromofluorobenzene	12.853	95	66567	44.634	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.260%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.389	101	33194	28.442	ug/l	96
12) 1,1-Dichloroethene	4.359	96	8811	7.710	ug/l	87
24) 1,1-Dichloroethane	6.583	63	3824	1.766	ug/l #	86
27) cis-1,2-Dichloroethene	7.500	96	4587	3.115	ug/l	91
30) Chloroform	7.977	83	3119	1.254	ug/l	92
38) Carbon Tetrachloride	8.365	117	3229	1.502	ug/l	89
44) Trichloroethene	9.359	130	2347129	1432.044	ug/l	100
55) 1,1,2-Trichloroethane	11.023	97	1314	0.836	ug/l #	94
64) Tetrachloroethene	11.106	164	4818	3.188	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
Data File : VN075446.D
Acq On : 28 Nov 2022 14:11
Operator : JC\MD
Sample : N5757-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE137D7-20221121

Quant Time: Nov 29 05:31:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

