

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
 Data File : VN075450.D
 Acq On : 28 Nov 2022 15:48
 Operator : JC\MD
 Sample : N5786-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 29 05:32:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	126111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	221357	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	192917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	76530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	41322	61.042	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 122.080%	
35) Dibromofluoromethane	8.171	113	48670	55.971	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.940%	
50) Toluene-d8	10.571	98	101566	51.859	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.720%	
62) 4-Bromofluorobenzene	12.853	95	76033	52.597	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.200%	
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	1637	1.030	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
Data File : VN075450.D
Acq On : 28 Nov 2022 15:48
Operator : JC\MD
Sample : N5786-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 29 05:32:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

