

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
 Data File : VN075452.D
 Acq On : 28 Nov 2022 16:35
 Operator : JC\MD
 Sample : N5786-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Nov 29 05:32:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	128112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	227292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	196447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	77446	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	42912	62.400	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	124.800%
35) Dibromofluoromethane	8.171	113	50266	56.297	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.600%
50) Toluene-d8	10.571	98	105352	52.388	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.780%
62) 4-Bromofluorobenzene	12.853	95	78482	52.873	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112822\
Data File : VN075452.D
Acq On : 28 Nov 2022 16:35
Operator : JC\MD
Sample : N5786-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Nov 29 05:32:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

